

PAC HOSPITAL KAMRA (DISTT ATTOCK)
TENDER NOTICE FOR ELECTRO MEDICAL EQUIPMENT
FINANCIAL YEAR: 2024-25

1. Quotations are invited from the sales tax registered firms for purchase of following Misc item according to the specifications mentioned against each for used in PAC Hospital Kamra:-

S No	Noun	Unit of Issue	Qty
a.	Vein Finder	Each	As required
b.	Infant Resuscitation Trolley (Warmer)	Each	01

❖ **SPECIFICATIONS OF VEIN FINDER:-**

1. Light Source : Near Infrared
2. Projection Resolution : Minimum 800 X 400
3. Projection Size : Very Small, Small, Medium & Large
4. Projection Brightness : 4 levels adjustable
5. Projection Colour : Green, Green Background, Black White, Red, Purple & Purple Background
6. Depth prompts : Deep, Medium & Shallow
7. Resolution speed should be minimum 50 frames per second.
8. Automatically goes into sleeping mode while not in use.
9. Battery working time minimum 2 hours.
10. Vein Finder could be used with mobile stand, desktop / bench top and handheld.
11. Mobile stand and desktop / bench top both are required with the equipment.
12. Two years warranty/ guarantee required of the equipment from date of installation.

"Or Equivalent"

❖ **SPECIFICATIONS OF INFANT RESUSCITATION TROLLEY (WARMER):-**

1. Pre-warm, Manual and Baby modes required.
2. Trolley and Baby temperature separately display.
3. Baby temperature control range : Min 34.5 °C Max 37.5 °C
4. Temperature of sensor display range : Min 5 °C Max 65 °C
5. Accuracy of skin temperature sensor should be less than or equivalent to 3 °C.
6. Temperature uniformity of mattress should be less than or equivalent to 2 °C.
7. Inclination of the bassinet having at least 05 positions.
8. Adjustable horizontal and Incline angle required.
9. Detachable panels required.
10. APGAR timer required.
11. Alarm indications required.
12. Two years warranty/ guarantee required of the equipment from date of installation.

"Or Equivalent"

10 pac 1411 - 212. pdf

STANDARD / SPECIAL CONDITION

- (a) Payment as per PAC Board purchase procedure.
- (b) Store on receipt will be accepted after inspection by EME staff of PAC Hospital.
- (d) Store must be latest production series (factory new).
- (e) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (f) Item will be accepted in accordance with standard specifications provided in the indent at consignee premises subject to acceptance by user and EME staff of PAC Hospital.
- (g) TO brochure & Technical manuals are needed to be provided by the supplier.
- (h) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (j) Two years Warranty / Guarantee required.
- (k) Delivery must be before 30-06-2025 after confirm / written purchase order / letter of intent.
- (l) As per standard of ISO 13485:2016.
- (m) Training on operation, maintenance and calibration of machine at PAC premises must be provided free of cost in warranty period i.e. 02 years as per Requirements / Technical manual of the EME.
- (n) At least ten years after sale spare supportability is to be provided.
- (p) OEM conformance and quality assurance certificate is required.
- (q) The repair / sales centre of the machine should be available at Rawalpindi / Islamabad / Peshawar or Attock.

2. Sealed Technical & Commercial quotations with the title "**Tender for Electro Medical Equipment of Vein Finder and Infant Resuscitation Trolley (Warmer)**" must be submitted to the undersigned on **02 December, 2024** up to **1030 Hrs** positively. **5% Bid Security** must be present in the sealed Commercial Quotation. The photocopy of **5% Bid Security** with black out value column should be present in the sealed Technical Quotation to ensure that original **5% Bid Security** is present in the sealed Commercial Quotation. Tenders would be opened at **1100 Hrs** on the same date.

3. Previous performance of the suppliers will be considered while accepting / rejecting the tender.

OFFICER COMMANDING
PAC Hospital Kamra (ATTOCK)
051-909902352