

**PAC HOSPITAL KAMRA (DISTT ATTOCK)**  
**TENDER NOTICE FOR ELECTRO MEDICAL EQUIPMENT**  
**FINANCIAL YEAR: 2024-25**

1. Quotations are invited from the sales tax registered firms for purchase of following Misc item according to the specifications mentioned against each for used in PAC Hospital Kamra:-

S No	Noun	Unit of Issue	Qty
a.	Dental X-Ray (Digital Imaging System)	Each	01

**SPECIFICATIONS:-**

❖ **Phosphors Plate System:-**

➤ Theoretical Resolution	:	20 lp/mm to 30 lp/mm
➤ Scan Time	:	1.5 sec to 3 sec
➤ Scan Time (HD mode)	:	2 sec to 4 sec
➤ Connection Ethernet	:	RJ-45
➤ Dimensions	:	Min L 154mm X D 204mm X H 193mm Max L 160mm X D 210mm X H 200mm
➤ Weight	:	2.5 Kg to 3 Kg
➤ Operating Voltage	:	100V to 240V
➤ Frequency	:	50Hz to 60Hz

❖ **Imaging Plates Sizes:-**

➤ Size 0	:	22mm X 35mm
➤ Size 1	:	24mm X 40mm
➤ Size 2	:	31mm X 41mm

➤ **Operating System:-**

➤ Operating System	:	Core i7 to Core i9
➤ RAM	:	16GB to 20GB
➤ ROM	:	1TB to 2TB
➤ Keyboard	:	Wireless
➤ Mouse	:	Wireless
➤ Display	:	20 to 24 inches LED

➤ **Power Back-Up:-**

➤ UPS 1KVA must be online.

**STANDARD / SPECIAL CONDITIONS**

- (a) Payment as per PAC Board purchase procedure.
- (b) As per standard of ISO 13485:2016.
- (c) Two years Warranty/Guarantee required.
- (d) Store on receipt will be accepted after inspection by EME staff of PAC Hospital
- (e) Store must be latest production series (factory new).
- (f) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.

- (g) Item will be accepted in accordance with standard specifications provided in the indent at consignee premises subject to acceptance by user and EME staff of hospital.
- (h) TO brochure & Technical manuals (02 Copies) are needed to be provided by the supplier.
- (j) Training on operation, maintenance and calibration of machine at PAC premises must be provided free of cost in warranty period i.e. 02 years as per Requirements / Technical manual of the EME.
- (k) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (l) At least ten years after sale spare supportability is to be provided.
- (m) OEM conformance and quality assurance certificate is required.
- (n) The repair / sales centre of the machine should be available at Rawalpindi / Islamabad / Peshawar or Attock.
- (p) Country of origin and manufacturer certification FDA, CE, MHLW any two (02) approved
- (q) Delivery must be before 30-06-2025.

2. Sealed Technical & Commercial quotations with the title "**Tender for Electro Medical Equipment of Dental X-Ray (Digital Imaging System)**" must be submitted to the undersigned on **07 January, 2025** up to **1030 Hrs. 5% Bid Security** must be present in the sealed Commercial Quotation. The photocopy of **5% Bid Security** with black out value column should be present in the sealed Technical Quotation to ensure that original **5% Bid Security** is present in the sealed Commercial Quotation. Tenders would be opened at **1100 Hrs** on the same date.

3. **Only one offer** can be made in relevant currency for same item and quoted price must be in **PKR**. Multiple rates, if quoted, will be rejected. Multiple quotations against the tender will also be rejected. Vendor must attach a **comparison sheet between specifications and special / standard conditions of PAC Hospital and your offered quote** with technical quotation.

3. Previous performance of the suppliers will be considered while accepting / rejecting the tender.

**Officer Commanding**  
PAC Hospital Kamra (Attock)  
051-909902352

*Pac 2012-71*