

PAC HOSPITAL KAMRA (DISTT ATTOCK)
TENDER NOTICE FOR ELECTRO MEDICAL EQUIPMENTS
FINANCIAL YEAR: 2024-25

1. Quotations are invited from the sales tax registered firms for purchase of following items according to the specifications mentioned against each for used in PAC Hospital Kamra:-

S No	Noun	Unit of Issue	Qty
01	Laparoscope (Telescope)	Each	01
02	Microscope	Each	01
03	Centrifuge Machine	Each	01

❖ **SPECIFICATIONS OF LAPAROSCOPE:-**

- Straight forward Telescope 0°
- Enlarged view
- Diameter : 10mm
- Length : 31cm
- Autoclaveable
- Fiber optic light transmission incorporated
- Color code : Green

❖ **SPECIFICATIONS OF MICROSCOPE:-**

- Binocular compound microscope
- Optical System : Universal identity system
- Head : binocular eye piece with 10x magnification, 360-degree rotating, -30-degree tilt, Inter-pupillary distance 45 – 75 mm.
- Revolving nosepiece fixed, front oriented.
- Stage holds single specimen with stage height adjustment mechanism
- Condenser : Abbe condenser with build-in aperture iris diaphragm.

❖ **SPECIFICATIONS OF CENTRIFUGE MACHINE:-**

- Bench top
- Microprocessor based control with LED display
- Stainless steel centrifuge bowl
- Brushless motor
- Safety lid interlock
- Dynamic / automatic brake and imbalance detector
- Speed range : 1000 – 5000 rpm or beyond
- Timer range : 0 – 30 min or beyond
- Rotor type : fixed angle
- Rotor capacity 24 or more positions for 5 – 20 ml tubes

2 pac 88 - 268. POF

STANDARD / SPECIAL CONDITION

- (a) Payment as per PAC Board purchase procedure.
- (b) Store on receipt will be accepted after inspection by EME staff of PAC Hospital.
- (d) Store must be latest production series (factory new).
- (e) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (f) Item will be accepted in accordance with standard specifications provided in the indent at consignee premises subject to acceptance by user and EME staff of PAC Hospital.
- (g) TO brochure & Technical manuals (02 Copies) are needed to be provided by the supplier.
- (h) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (j) One year Warranty required.
- (k) Delivery within 02 weeks after confirm / written purchase order / letter of intent.
- (l) As per FDA standard of Europe / USA.

2. Sealed Technical & Commercial quotations with the title "**Tender for Electro Medical Equipment**" must be submitted to the undersigned on **26 August, 2024** up to **1030 Hrs** positively. **5% Bid Security** must be present in the sealed Commercial Quotation. The photocopy of **5% Bid Security** with black out value column should be present in the sealed Technical Quotation to ensure that original 5% Bid Security is present in the sealed Commercial Quotation. Tenders would be opened at **1100 Hrs** on the same date.

3. Previous performance of the suppliers will be considered while accepting / rejecting the tender.

OFFICER COMMANDING
PAC Hospital Kamra (ATTOCK)
051-909902352