

Ref: 1/SZH/Gyn-OT/2021/04

Date:04-02-2022

Mr. Rizwan Mahmood, Deputy Director, PPRA, Sector G-5/2, Islamabad. Phone# 051-9205728

Subject: -

TECHNICAL EVALUATION REPORT FOR THE TENDER TS461242E, I.E. TWO OPERATION THEATER TABLES, FOR GYNAECOLOGY DEPARTMENT SHAIKH ZAYED HOSPITAL, LAHORE,

Please find enclosed here with Technical Evaluation Report of said Medical Equipment of "Gynecology Department Shaikh Zayed Hospital, Lahore, for PPRA website publication.

After the publication, please send us the invoice, we shall return back with the pay order of that amount within couple of days.

Admin officer, 7/21

Shaikh Zayed Hospital, Lahore.

TECHNICAL EVALUATION REPORT

FOR TWO MOBILE OPERATION THEATER TABLES SHAIKH ZAYED HOSPITAL, LAHORE.

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore

2. Method of Procurement: Single Stage Two Envelops

3. Title of Procurement: Medical Equipment for shaikh Zayed Hospital, Lahore.

4. Tender Inquiry No.: Ref: 1/SZH/Gyn-OT/2021/02

5. PPRA Ref. No. (TSE): TS461242E

6. Date & Time of Bid Closing: 06/11/2021 at 11:30:00 AM

7. Date & Time of Bid Opening: 06/11/2021 at 12:00:00 Noon

No of Bids Received: 0

9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding

documents / feedback of the End User.

10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase

Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory

Committee.

	Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Ver	rtex Medical	Qualified	Accepted as per Criteria of bidding documents and final recommendation of the Ende User
M/S Spo Pvt.	ectrum Technologies	Not-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, (company neither provided accessories number nor catalogues, lateral tilt is only 15° which is not sufficient. Foot paddle is not there, leg portion are not split.)
M/S Eas Service	stern Technology es	Not-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, (The company has provided operation theatre tables in liver transplant unit (MEDIFIA, Germany MO00 603200) in 2017 which have very poor quality and batteries were replaced in December 2020. Backup services were also very poor as communicated by biomedical department
M/S G	M/S G Med Pvt. Ltd		Accepted as per Criteria of bidding documents and final recommendation of the Ende User
M/S Tot	M/S Total Technologies		Rejected as per Criteria of bidding Documents and recommendations of end user, (absence of foot paddle)

Technical Committee recommended M/S Vertex Medical and G Med Pvt. Ltd are responsive bidders and recommended for financial bid opening.

Signature:

Official Stanghaikh Zayed Hospital

ahore