



COMSATS University Islamabad (CUI), Lahore Campus

Defence Road, Off Raiwind Road, Lahore

Tel: 042-111-001-007 (Ext: 875), Fax: 042-99203100

Evaluation Report

(As per Rule 35 of Public Procurement Rules, 2004)

1. Name of Procuring Agency: COMSATS University Islamabad (CUI), Lahore Campus
2. Method of Procurement: Single Stage Two Envelope
3. Title of Procurement: Oxygen Index Analyzer for Department of Chemical Engineering under Swedish Funded Project awarded to Dr. Asad Ullah Khan (Case # 3402)
4. Tender Inquiry No: CUI-LHR-TN-21-1512
5. PPRA Ref No. (TSE): TS460214E
6. Date & Time of Bid Closing: 1400 hours on October 06, 2021
7. Date & Time of Bid Opening: 1430 hours on October 06, 2021
8. No. of Bids Received: 04 Nos.
9. Criteria for Bid Evaluation: Attached
10. Detail of Bid (s) Evaluation: Attached

Sr. No	Equipment Name	Company (M/s.)	Total Evaluated Cost (Rs.)	Comments
1	Oxygen Index Analyzer	Linker Enterprises	1,430,000/-	Bid accepted being technically qualified & financially lowest (Quoted Model: F101-D Qinsun Instrument Co Ltd China)
		Arshad Amjad & Abid (Private) Limited		Not Recommended due to followings: LCD Display dimensions are not as per BOQs 2. USB Ethernet port is not available along with the desktop device 3. Sample dimension size specifications are not as per BOQs 4. Glass column is not as per BOQs (i.e., Botrosilicate glass column) 6. No availability of Oxygen gas with cylinder as mentioned in the information supplied by the vendor. (Quoted Model : LABFAC LR-5 Automatic Oxygen Index Tester)

al-han
25/3/2022

MUHAMMAD SHAHID
Sr. Coordination Officer
Purchase Section
COMSATS University Islamabad,
Lahore Campus




COMSATS University Islamabad (CUI), Lahore Campus

Defence Road, Off Raiwind Road, Lahore

Tel: 042-111-001-007 (Ext: 875), Fax: 042-99203100

		Emerging Technologies	-	Not Recommended due to followings: 1. conflicting model numbers are mentioned by the vendor (Ttech-GBT2406-1) (FTech-GB/T 2406B) 2. LCD Display dimensions are not as per BOQs 3. USB Ethernet port is not available along with the desktop device 4. Glass column is not as per BOQs (i.e. Botrosilicate glass column) (Quoted Model: Ftech-GB/T 2406 B)
		OTC	-	Bid Rejected due to not mentioned the Quoted Model / Brand on our prescribed tender documents / BoQs.


(Muhammad Shahid) 25/03/2022
SCO (Purchase)

MUHAMMAD SHAHID
Sr. Coordination Officer
Purchase Section
COMSATS University Islamabad,
Lahore Campus