

TECHNICAL EVALUATION REPORT

FOR NCS / EMG / EP MACHINE – NEUROLOGY DEPARTMENT

SHAIKH ZAYED HOSPITAL, LAHORE.


1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Invitation to Bid, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/Neuro-Gynae-Eq/2022/03
5. PPRA Ref. No. (TSE): TS498342E
6. Date & Time of Bid Closing: 19/01/2023 at 11:00:00 AM
7. Date & Time of Bid Opening: 19/01/2023 at 11:30:00 AM
8. No of Bids Received: 03
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

| Name of Bidder | Technical Status (if applicable) | Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004. |
|--|-------------------------------------|---|
| M/S DIGIONICS PVT, LTD | Qualified | Accepted as per Criteria of bidding documents and final recommendation of the End User. |
| M/S MEDEQUIPS PVT, LTD. | Qualified | Accepted as per Criteria of bidding documents and final recommendation of the End User. |
| M/S SHIRAZI TRADING COMPANY PVT, LTD. | Qualified | Accepted as per Criteria of bidding documents and final recommendation of the End User. |

COMMITTEE RECOMMENDATIONS

Subsequently getting clarifications from **M/S Digionics Pvt Ltd & M/S Medequips Pvt Ltd**, Technical Evaluation Committee recommended that **M/S Digionics Pvt Ltd, M/S Medequips Pvt Ltd and M/S Shirazi Trading Company Pvt, Ltd** are technically responsive bidders, and are recommended for financial bid opening / purchase of "**NCS/EMG/EP MACHINE**" for Neurology Department.

Signature:


 EXECUTIVE DIRECTOR
 Shaikh Zayed Medical Complex,
 Lahore

Official Stamp:



TECHNICAL EVALUATION REPORT

FOR MOBILE OPERATING TABLES – GYNECOLOGY DEPARTMENT

SHAIKH ZAYED HOSPITAL, LAHORE.

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Invitation to Bid, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/Neuro-Gynae-Eq/2022/03
5. PPRA Ref. No. (TSE): TS498342E
6. Date & Time of Bid Closing: 19/01/2023 at 11:00:00 AM
7. Date & Time of Bid Opening: 19/01/2023 at 11:30:00 AM
8. No of Bids Received: 05
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

| Name of Bidder | Technical Status (if applicable) | Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004. |
|---------------------------------|-------------------------------------|---|
| M/S SHIRAZI TRADING | Qualified | Accepted as per Criteria of bidding documents and final recommendation of the End User. |
| M/S G MED PVT, LTD. | Not-Qualified | Rejected as per Criteria of bidding Documents and recommendations of end user, (<u>company Offered Model FAMED - SU 03, manufactured by Poland. G-med is not qualified as movements of leg and head section is only manual not electrical. Not meet Specification provided. So this firm is technically disqualified.</u>) |
| M/S EASTERN TECHNOLOGY SERVICES | Qualified | Accepted as per Criteria of bidding documents and final recommendation of the End User. |
| M/S DIGIONICS PVT, LTD. | Qualified | Accepted as per Criteria of bidding documents and final recommendation of the End User. |
| M/S VERTEX MEDICAL | Not-Qualified | Rejected as per Criteria of bidding Documents and recommendations of end user, (<u>company offered Model ETIUDA 4-02, manufactured by Alvo Medical, Poland is not qualified as they quoted the table, with weight bearing capacity of 275kg and we demanded 300kg as better. So this firm is technically disqualified.</u>) |

COMMITTEE RECOMMENDATIONS

TEC Committee recommended that **M/S Digionics Pvt Ltd, M/S Eastern Technology Services and M/S Shirazi Trading Company Pvt, Ltd** are technically responsive bidders, and recommended for financial bid opening / purchase of **“MOBILE OPERATING TABLE”** for Gynecology Department.

Signature: _____

Official Stamp: _____

EXECUTIVE DIRECTOR
Shaikh Zayed Medical Complex,
Lahore

[Handwritten Signature]
28-3-2023