

TECHNICAL EVALUATION REPORT

(As Per Rule 35 of PP Rules, 2004)

1. Name of Procuring Agency: Shaikh Zayed Postgraduate Medical Institute, Lahore.
2. Method of Procurement: Single Stage Two Envelops.
3. Title of Procurement: Up Gradation of Radiology Department at Shaikh Zayed Postgraduate Medical Institute, Lahore. PHASE – III.
4. Tender Inquiry No.: 2284
5. PPRA Ref. No. (TSE): TS532142E.
6. Date & Time of Bid Closing: 29-02-2024, 11:30AM.
7. Date & Time of Bid Opening: 29-02-2024, 12:00PM
8. No of Bids Received: 04.
9. Criteria for Bid Evaluation: Knock Down.
10. Details of Bid(s) Evaluation:

Equipment Name	Name of Bidder	Technical Marks (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
<u>Mobile DR</u>	1. M/S Hoorra Pharma Pvt Ltd		Technically Accepted as meet the given criteria/ Bidding requirement
	2. M/S Shahco Medical Pvt Ltd		Rejected on the grounds: <ul style="list-style-type: none">• CDR not attached as per clause 13.2 page # 20 of SBD.• Exclusive Authorization / Sole Agent Certificate is not attached.• Firm does not have 5-year experience as per requirement IBT 3.8 and no satisfactory performance report is attached.
Mobile X-ray	1. M/S Hoorra Pharma Pvt Ltd		Technically Accepted as meet the given criteria/ Bidding requirement
	2. M/S Shahco Medical Pvt Ltd		Rejected on the grounds: <ul style="list-style-type: none">• CDR not attached as per clause 13.2 page # 20 of SBD.• Exclusive Authorization / Sole Agent Certificate is not attached.• Firm does not have 5-year experience as per requirement IBT 3.8 and no satisfactory performance report is attached.
DR System	1. M/S Hoorra Pharma Pvt Ltd		Technically Accepted as meet the given criteria/ Bidding requirement
	2. M/S Radiant Medical Pvt Ltd		Rejected on the grounds:


Accepted

Saukat

		<ul style="list-style-type: none"> • Exclusive Authorization / Sole Agent Certificate is not attached. • Attached FDA is not of quoted model Dryview 6950. • Non-Motorized unit and Non-Motorized chest stand are not offered which are mandatory to quote in the bidding documents. • X-ray tube and collimator are not from OEM. • Country of manufacture as china.
	3. M/S Medical Equipment & Systems	Rejected on the grounds: CDR not attached as per clause 13.2 page # 20 of SBD.
Color Doppler low end	1. M/S Hoorra Pharma Pvt Ltd	Rejected due to Country of manufactured of quoted model as Korea which is contradict to Corrigendum.
	2. M/S Medical Equipment & Systems	Rejected on the grounds: CDR not attached as per clause 13.2 page # 20 of SBD.
	3. M/S Shirazi Trading Co. Pvt Ltd	<p>Rejected on the grounds:</p> <ul style="list-style-type: none"> • Exclusive Authorization / Sole Agent Certificate is not attached • Attached brochure is pertains to Arietta-50LE whereas as quoted model is Arietta 50. • Performance certificates are attached of other models.
Digital Mini C-arm	1. M/S Shahco Medical Pvt Ltd	<p>Rejected due to</p> <ul style="list-style-type: none"> • CDR not attached as per clause 13.2 page # 20 of SBD. • Vendor Past performance of quoted Product is not attached

11. Any other additional / supporting information, the procuring agency may like to share.

Saulat
Prof. Dr. Saulat Sarfraz,
 Head of Radiology Department
 Shaikh Zayed Hospital,
 Lahore


Dr. Muhammad Afaaq Agha
 Deputy Administrator (Pharmacy)
 Shaikh Zayed Hospital Lhr.