



# COMSATS University Islamabad (CUI), Lahore Campus

Defence Road, Off Raiwind Road, Lahore

Tel: 042-111-001-007 (Ext: 875), Fax: 042-99203100

## Evaluation Report

(As per Rule 35 of Public Procurement Rules, 2004)

1. Name of Procuring Agency: COMSATS University Islamabad (CUI), Lahore Campus
2. Method of Procurement: Single Stage Two Envelope
3. Title of Procurement: Thermo Gravimetric Analyzer (TGA) under CATBM PC-1 (Case # 3398)
4. Tender Inquiry No: CUI-LHR-TN-21-1503
5. PPRA Ref No. (TSE): TS458558E
6. Date & Time of Bid Closing: 1400 hours on September 15, 2021
7. Date & Time of Bid Opening: 1430 hours on September 15, 2021
8. No. of Bids Received: 09 Nos.
9. Criteria for Bid Evaluation: Attached
10. Detail of Bid (s) Evaluation: Attached

Sr. No	Equipment Name	Company (M/s.)	Total Evaluated Cost (Rs.)	Comments
1	Thermo Gravimetric Analyzer (TGA)	Linker Enterprises	3,976,800	<b>Bid accepted being technically qualified &amp; financially lowest</b> (Model: BXT-DSC-TGA-1250 Glomro Industrial Co Ltd with UPS and Dell Desktop System with 24'' screen)
		United Traders	4,991,664.6	<b>Bid accepted being technically qualified.</b> (Model: TR-TCA 101 China Trytec Technology included Dell Optiplex 7040 Ci5, Hp LJ Pro M15w, Nitrogen China)
		Systec Technologies	4,550,000	<b>Bid accepted being technically qualified.</b> (Model: TGA 101 Labmen Instrument China)
		Arshad Amjad & Abid (Private) Limited	5,904,349	<b>Bid accepted being technically qualified.</b> (Model: TGA 101 Lab Kits)




# COMSATS University Islamabad (CUI), Lahore Campus

Defence Road, Off Raiwind Road, Lahore

Tel: 042-111-001-007 (Ext: 875), Fax: 042-99203100

	<b>Kamstec International</b>	9,945,000	<b>Bid accepted being technically qualified.</b> (Model: TA Instruments USA Discovery TGA 55)	
	<b>Nexus Corporation</b>	-	<b>Not Recommended by Technical Committee</b> (Model: TGA1001 Xiang Yi China)	
			<b>Required</b> Resolution: 0.1 u g or better	<b>Offered</b> (No Proof from OEM was provided in technical bid and subsequent query)
	<b>Emerging Technologies</b>	-	<b>Not Recommended by Technical Committee</b> (Model: BXT-TGA1150 Shanghai Glomao Industrial Co. Ltd)	
			<b>Required</b> Resolution: 0.1 u g or better	<b>Offered</b> 0.01 mg , (BXT-TGA-1150 (shanghai Glomro)
	<b>Innovative Power Engineering</b>	-	<b>Not Recommended by Technical Committee</b> (Model: SKZ 1053 SKZ Industrial China)	
			<b>Required</b> Resolution: 0.1 u g or better	<b>Offered</b> 0.1 mg
	<b>Worldwide Scientific</b>	-	<b>Not Recommended by Technical Committee</b> (Model: BK-TGA 1000B, Biobase)	
			<b>Required</b> Resolution:0.1 u g or better	<b>Offered</b> 0.01 mg,

  
 (Muhammad Shahid) 04/04/2022  
 SCO (Purchase)  
**MUHAMMAD SHAHID**  
 Sr. Coordination Officer  
 Purchase Section  
 COMSATS University Islamabad,  
 Lahore Campus