

## **Evaluation Report**

[As per Rule 35 of PPRA Rules-2004]

1. Name of Procuring Agency

PIAC

2. Method of Procurement

Single Stage Two Sealed Envelope Basis.

3. Title of Procurement

PPE Suits (Isolation Gown)

4. Tender Inquiry No.

FSD-018-20 (Tender No 02)

5. PPRA Ref No.

TS 423 454 E

6. Date & Time of Bid Closing

22-06-2020 1030 Hrs

7. Date & Time of Bid Opening

22-06-2020 1030 Hrs

8. No of Bids Received

SEVEN

9. Criteria for Bid Evaluation

As per specification attached with tender documents.

10. Details of Bid Evaluation : User Department/Inspection & Quality Control.

|   | The state of the s |                       |                                  |  |
|---|--|-----------------------|----------------------------------|--|
| Bidder's Name   | Mark Independent Lab M/s. TTI Report Nos Dt; 23-06-2020  | Evaluation<br>Results | Evaluated<br>Cost /<br>Financial | Rule / Regulation / SBD* / Policy /<br>Basis for Rejection / Acceptance<br>as per Rule-35 off PP Rules, 2004.        |
| 01/07 – HUNAN Yourfield Medical<br>Devices Co., Ltd China | 10448-20   | REJECTED              | Not Opened                       | Flame Resistance of Textile does not comply.   |
| 02/07 - DALIEN International<br>Limited Hong Kong         | 10449-20   | REJECTED              | Not Opened                       | Flame Resistance of Textile does not comply.   |
| 03/07 – AL-AYUB Packages (Pvt)<br>Ltd – Karachi.          | 10450-20   | REJECTED              | Not Opened                       | Structure & Stitching not<br>comply. Water Resistance<br>(Penetration and Hydrostatic<br>Resistance) does not Comply |
| 04/07 – KOLACHEE International (Pvt) Limited              | 10451-20   | REJECTED              | Not Opened                       | Flame Resistance of Textile does not comply.   |
| 05/07 – MS Enterprises Karachi                            | 10452-20   | REJECTED              | Not Opened                       | Structure and Stitching does not comply.   |
| 06/07 – WINNER Medical Co. Ltd<br>China                   | 10453-20   | RECOMMENDED           | US\$ 99,500.00                   | Comply with all Test Results completely.   |
| 07/07 – THE EXPERTS Global Supplies (Pvt) Ltd.            | 10454-20   | REJECTED              | Not Opened                       | Flame Resistance of Textile does not comply.   |

Lowest Evaluated Bidder: M/s. WINNER Medical Co. Ltd China Any other Additional / supporting information, the procuring agency may like to share.

Note → Tender Samples Laboratory Test were performed through independent Laboratory

- M/s. TTI Testing Laboratories
- Address; 347-S, Quaid-E-Azam Industrial Estate, Kot Lakhpat, Lahore 54770, Pakistan.
- Attention : Mr. Ali Ashraf (GM Operations)
- UAN 111-786-001 Ext: 108 | Cell: +92 320 8484540 Email: ashraf@ttilabs.net | URL: www.tttilabs.net

Muhammad Saleem Akhtar

Alalan 127 ron

**DGM Purchases Commercial** 

\*Standard Bidding Documents (SBD)

Purchases Flight Kitchen Section,

Email: zameer.hussain@piac.aero / dgmpc@piac.aero